

Guide Wire [INSTRUCTION FOR USE]

1. INTENDED PURPOSE

The product play a support and guide the role when used with “J” catheter and other medical devices in endoscope by professional physician.

2. CONTRAINDICATION

No absolute contraindications when used correctly by medical personnel.

3. HANDING AND STORAGE

Stored in the relative humidity of not more than 80%, non-corrosive gases, a common, dry, well-ventilated and clean environment.

4. STRUCTURE AND SPECIFICATIONS

The product is consist of core wire, coil, clasp, tip protector, booster.

Specifications:

model	diameter	Tip	PTFE Coated	length	
BR-GW	0.018	S/J	/	45、60、80、120、150、180、260	
	0.021			45、60、80、120、150、180、260	
	0.025			45、60、80、120、150、180、260	
	0.032			45、60、80、120、150、180、260	
	0.035			45、60、80、120、150、180、260	
	0.038			45、60、80、120、150、180、260	
BR-GWC	0.018		S/J	PTFE	45、60、80、120、150、180、260
	0.021				45、60、80、120、150、180、260
	0.025				45、60、80、120、150、180、260
	0.032				45、60、80、120、150、180、260
	0.035				45、60、80、120、150、180、260
	0.038				45、60、80、120、150、180、260
model	diameter	Tip		PTFE Coated	length
BR-NGW	0.018	S/J		/	45、60、80、120、150、180、260
	0.021				45、60、80、120、150、180、260
	0.025				45、60、80、120、150、180、260
	0.032				45、60、80、120、150、180、260
	0.035				45、60、80、120、150、180、260
	0.038		45、60、80、120、150、180、260		
BR-NGW C	0.018		S/J	PTFE	45、60、80、120、150、180、260
	0.021				45、60、80、120、150、180、260
	0.025				45、60、80、120、150、180、260
	0.032				45、60、80、120、150、180、260
	0.035				45、60、80、120、150、180、260
	0.038				45、60、80、120、150、180、260

5. PRODUCT PERFORMANCE

The product has good maneuverability, flexibility, visibility, traceability, smoothness and support.

6. INSTRUCTION FOR USE

1. Check the packaging before use. It is strictly prohibited if damaged.
2. Tear off the dialysis paper and remove it together with the guide wire sheath.
3. Remove the guide wire sheath and push the 3-5cm guide wire by hand to check its strength and softness.
4. The guide wire booster is used to send the guide wire through the puncture needle into the artery and vein and push it to the desired depth. The "J" type head may be slightly turned to enter.

WARNING: Do not pull the guide wire at an angle from the puncture needle to avoid serious damage to the guide wire.

5. Grip the guide wire, remove the puncture needle.

NOTE: Always hold the guide wire at all times. Adjust the depth of guide wire according to the centimeter mark on the guide wire. The centimeter mark on the guide wire is measured from the "J" type head, a bar means 10 cm, two bars means 20 cm, three bars means 30 cm and so on.

6. Hold the catheter and draw the guide wire when the catheter reaches a desired depth.
7. Check whether the guide wire is intact after drawing out.

7. PRODUCT LABEL AND PACKAGING (See the label or packaging)

1. Product name, model and specification.
2. Production date, batch lot.
3. Product license, registration number, product technology requirement.
4. Sterilization expiry date.

8. CAUTION

1. Procedure must be performed by trained medical personnel.
2. Sterilization is valid for three years, please confirm date of sterilization before use, do not use if expiry.
3. Do not use if luer taper leakage.
4. The device is designed and intended for SINGLE USE ONLY.
5. Do not resterilize and/or reuse, otherwise patients could get the risk of infection.
6. Sterilized by ethylene oxide. Do not use if the package damaged.



Caution



Use by date



Keep dry



Keep away from sunlight

REF

Catalogue number



Date of manufacture



Manufacturer



Do not use




Sterilized Using EO



Do not use if package is damaged



Batch code



BrosMed Medical B.V.

ADDRESS: Mgr. Buckxstraat 8, 6134 AP Sittard, the Netherlands

TEL: +31 850140901

Mail: cs@brosmed.com



Brun Medical Co., Ltd.

ADDRESS: Building B8, Entrepreneurship and Innovation Park, 88 Changtang Road,
Yunlong Demonstration Zone, Zhuzhou 412000 HUNAN, CHINA

TEL: +0731-22029666

WEB: www.brunmd.com